



AccuTrak™

aTPO Calibration Verification Panel

INTENDED USE

AccuTrak aTPO Calibration Verification Panel solutions are intended for *in vitro* diagnostic use in the determination of calibration verification and verification of reportable range in automated instrument systems for the following analyte: Anti-Thyroid Peroxidase (aTPO).

AccuTrak Panels are non-automated and intended for laboratory professional use only.

Each panel consists of one bottle each of Levels 1 through 5. Each bottle contains 3.0 milliliters.

SUMMARY

The AccuTrak Calibration Verification Panel will assist in the documentation of calibration verification and verification of reportable range required by many inspection agencies. The solutions will also provide assistance when troubleshooting instrument systems, reagent problems and calibration anomalies.

REAGENTS

Reactive Ingredients:

Purified chemicals and constituents of human and/or animal source in human serum.

Nonreactive Ingredients:

Preservatives and stabilizers.

Precautions and Warnings:

For In Vitro Diagnostic Use

Disposal of all waste material should be in accordance with local guidelines.

Full SDS is available at www.mainestandards.com/doc-search.

Consult this SDS for detailed safety information including handling spills.

WARNING: Potentially Biohazardous

Human source material is considered potentially biohazardous. Use the Centers for Disease Control (CDC) recommended universal precautions for handling AccuTrak products.

STORAGE AND STABILITY

The AccuTrak aTPO Calibration Verification Panel is stored at -10° to -25°C. **Do NOT store in a frost-free freezer.** Panels are stable until the expiration date printed on the storage container when handled according to instructions. Do not use the Kit past its expiration date. **A maximum of four (4) freeze-thaw cycles is recommended.**

PREPARATION

Prior to use, remove the AccuTrak aTPO Calibration Verification Panel from storage and allow to come to room temperature (18° to 25°C). Invert gently several times before dispensing.

To maximize stability, it is recommended that exposure to room temperature be minimized. Tightly cap opened bottles and return to -10° to -25°C immediately after dispensing.

Discard any solutions that appear to have gross bacterial contamination.

AccuTrak Calibration Verification Panels should be treated in the same manner as patient samples. If dilutions or pre-treatment are required as part of the testing procedure, follow the manufacturer's instructions.

ASSAY

Analyze each level in replicates.

LIMITATIONS

AccuTrak Calibration Verification Panel solutions are not intended for use as routine quality control materials or as calibration materials.

APPROXIMATE VALUES

The values listed are approximate and provided only for reference. Actual observed values will vary depending on the lot number, temperature, reagent and/or calibrator.

Level	Range	Units
1	9.3 – 26.7	IU/mL
2	135 – 165	IU/mL
3	253 – 314	IU/mL
4	373 – 459	IU/mL
5	498 – 598	IU/mL

The following symbols may be used where applicable in labeling for LGC Clinical Diagnostics products:

-  Lot Number
-  Expiration Date
-  Manufacturer
-  Temperature Limit
-  In Vitro Diagnostic Medical Device
-  Catalog Number
-  Instructions for Use
-  Biological Risk
-  Do Not Reuse
-  Caution

ORDERING INFORMATION

ORDER NO.: 2400-0254

AccuTrak aTPO

Calibration Verification Panel:
5 x 3 mL

Any serious incident occurring in relation to the use of this device shall be reported to the Manufacturer as well as the Competent Authority of the member state in which the user/patient is established.

CONTACT INFORMATION:

 LGC Clinical Diagnostics, Inc.
221 US Route 1
Cumberland Foreside, ME 04110 USA

1-800-377-9684
+1-207-892-1300
+1-207-892-2266 Fax
msc.sales@LGCGroup.com
msc.techsupport@LGCGroup.com
www.mainestandards.com

Issue Date	Modification
Jan 9, 2025	Initial Issue